

Declaration of Conformity

(in accordance with ISO/IEC 17050-1)

Company Information:

American Orthodontics
3524 Washington Avenue
Sheboygan, WI 53081

Certificate Number: 201603 -12a

Certificate Issue Date: 03/07/16

Certificate Expiry Date: 11/22/17

This certificate is valid until Expiry Date listed above.

Medical Device Detail:

Product/Family Category:

Adhesives, Dental Impression

Classification: *Classification is according to Directive 93/42/EEC, Annex IX*

Class *IIa per Rule 5, 3rd indent*

Product(s) Tradename(s):

Eagle FluorSure™
Eagle MTP™ (Moisture Tolerant Primer)
No Mix:30™
Bandtite™

System(s)/Configuration(s) associated with Product(s) Tradename(s):

Eagle MTP™ (Moisture Tolerant Primer)
No Mix:30™ (Syringe Kit, Jar Kit)
No Mix:30™ - Activator/Sealant
Bandtite™ Glass Ionomer Self Cure Band Cement
Bandtite™ Band Cement

CE Authorized Representative:

Emergo Europe
Molenstraat 15
2513 BH The Hague, The Netherlands

Assessment Details:

Notified Body: TUV Austria
CE Certificate #: TUV-A-MT-1/13/QE011
Issue Date: 05/01/13

Route of Directive: Annex VI *of the Directive 93/42/EEC on Medical Devices.*

Standards Applied:

Harmonized Standards: IEC ISO 62336, EN ISO 14971, EN 1041, EN ISO 10993-1, 10993-3, 10993-5, 10993-10, 10993-11

Other Standards: ISO 20795-1, ISO 7405, ISO/TS 11405

Declaration statement:

Each kind of medical device to which the technical documentation applies complies with the applicable provisions of the essential principles and the classifications rules before being shipped. The above products are in compliance with Annex I Essential Requirements of Directive 93/42/EEC. The product is CE marked.

Signed for and on behalf of:

American Orthodontics
Sheboygan, WI

Company Representative:


Trang Adams / Regulatory Affairs Specialist

